



EUROPEAN INSPECTION AND CERTIFICATION COMPANY S.A.

CERTIFICATE No.: CN.CE.0652-10/10

CERTIFICATE OF FINAL ASSESSMENT MONITORING
PRESSURE EQUIPMENT DIRECTIVE 97/23/EC (MODULE A1)


Manufacturer's Name : D MASS GmbH
Manufacturer's Address : Lurgiallee 6-8 D-60439 Frankfurt/Main
Type of Pressure Equipment : PRESSURE GAUGES (MBB,MBS)
Category : II
No. of Samples assessed : 4

It is hereby certified, that EUROCERT S.A. has assessed the above pressure equipment against the provisions of Annex III (Module A1) and that the manufacturer is authorized to affix on this pressure equipment the CE Mark and Eurocert's Identification No. as displayed below:

CE 1128

This certificate is subject to unexpected surveillance visits by Eurocert SA.

ATHENS, 07-10-2010
For EUROCERT


GEORGE SIFONIOS
DIRECTOR OF DEVELOPMENT

THE DATA OF THIS CERTIFICATE WERE GATHERED WITH EVERY POSSIBLE THOROUGHNESS
THIS CERTIFICATE REFLECTS THE FINDINGS OF THE TIME AND PLACE OF THE AUDIT
REPRODUCTION OF THIS DOCUMENT IS STRICTLY FORBIDDEN



Product Certification
Cert. No. 21
AT13.18/F18/01.04-2009

89, CHLOIS STR. & LIKOVRISEOS, 144 52 METAMORFOSI, ATHENS, GREECE
Tel.: ++30 210 62.52.495, 30 210 62.53.927 - Fax: ++30 210 62.03.018
Internet site: www.eurocert.gr - e-mail: eurocert@otenet.gr





EUROPEAN INSPECTION AND CERTIFICATION COMPANY S.A.

CERTIFICATE No.: CN.CE.0652-09/10

TECHNICAL DESCRIPTION
BOURBON TUBE PRESSURE GAUGE
DMASS GmbH

Technical Characteristics of Pressure Equipment:

Description of the Equipment		BOURBON TUBE PRESSURE GAUGE			
No.	Model / Type	Nominal Size	Accuracy	Measurement Range	Working Temperature
1	MBB	NS 50	2.5	'0 to 1 bar' up to '0 to 40 bar'	+60 °C maximum
2			1.6		
3		NS 63	1.6	'0 to 0.6 bar' up to '0 to 600 bar'	
4		NS 100	1.0	'0 to 0.6 bar' up to '0 to 600 bar'	
5	MBS	NS 50	2.5	'0 to 1 bar' up to '0 to 40 bar'	+200 °C maximum without liquid filling +100 °C maximum with liquid filling
6			1.6		
7		NS 63	1.6	'0 to 0.6 bar' up to '0 to 40 bar'	
8		NS 100	1.0	'0 to 0.6 bar' up to '0 to 60 bar'	



Product Certification
Cert. No. 21
ΔΠ13.18/E18/21-04-2009

89, CHLOIS STR. & LIKOVRISEOS, 144 52 METAMORFOSI, ATHENS, GREECE
Tel.: ++30 210 62.52.495, 30 210 62.53.927 - Fax: ++30 210 62.03.018
Internet site: www.eurocert.gr - e-mail: eurocert@otenet.gr

